



CERTIFICATE OF ANALYSIS

PRODUCT CT0693B
CLR15 CLARITHROMYCIN

LOT NUMBER 3729101

EXPIRY DATE 2026.09.26

DATE OF MANUFACTURE 2023.09.27

Delivery/Customer information

Date Printed

2023.10.05

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 14.0

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	27	26 - 32
<i>Haemophilus influenzae</i> ATCC®49247	13	11 - 17
<i>Streptococcus pneumoniae</i> ATCC®49619	29	25 - 31

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.10.05

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096