

CERTIFICATE OF ANALYSIS

PRODUCT CT1628B

TZP40 PIPERACILLIN/TAZOBACTAM

LOT NUMBER 1351450

EXPIRY DATE 2016.06.30

DATE OF

MANUFACTURE 2013.05.22

Delivery/Customer information

Date Printed 2022.11.23

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc)	PRL30	33.1
Assay (mcg/disc)	YTR10	12.1

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits
		(mm)
Staphylococcus aureus ATCC®25923	30	23 - 30
Escherichia coli ATCC®25922	23	23 - 26
Pseudomonas aeruginosa ATCC®27853	26	25 - 29

Quality control ranges are internally derived based on historical data.

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.

Lot Accepted. 2013.05.29

Carol MacBride

Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

