

CERTIFICATE OF ANALYSIS

PRODUCT CT1633B
MXF5 MOXIFLOXACIN

LOT NUMBER 6230144

EXPIRY DATE 2027.02.10

DATE OF MANUFACTURE 2025.02.10

Delivery/Customer information	
Date Printed	2025.02.21
Delivery No.	
Customer	
Customer Order Number	

Microbiological Performance

Assay (mcg/disc) 5.19

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	29	28 - 35
<i>Escherichia coli</i> ATCC®25922	28	28 - 35
<i>Pseudomonas aeruginosa</i> ATCC®27853	20	17 - 25
<i>Haemophilus influenzae</i> ATCC®49247	32	31 - 39
<i>Streptococcus pneumoniae</i> ATCC®49619	27	25 - 31

Tested in accordance with current EUCAST methodology

<i>Escherichia coli</i> ATCC®25922	28	28 - 35
<i>Staphylococcus aureus</i> ATCC®29213	26	25 - 31
<i>Streptococcus pneumoniae</i> ATCC®49619	29	24 - 30
<i>Haemophilus influenzae</i> ATCC®49766	31	30 - 36

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.02.21

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.