

CERTIFICATE OF ANALYSIS

PRODUCT CT1727B
SCF105 SULBAC/CEFOPERAZONE

LOT NUMBER 6248338

EXPIRY DATE 2028.03.26

DATE OF MANUFACTURE 2025.03.27

Delivery/Customer information

Date Printed

2025.04.03

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) : Sulbactam 36.2
Assay (mcg/disc) : Cefoperazone 82.0

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	26	24 - 30
<i>Escherichia coli</i> ATCC®25922	27	27 - 33
<i>Pseudomonas aeruginosa</i> ATCC®27853	25	22 - 28
<i>Acinetobacter calcoaceticus</i> ATCC®43498	30	26 - 32

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.03

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com