

G & M Procter Ltd. Certificate of Analysis

PRODUCT **PO0839A**
CAMPYLOBACTER C.A.T. AGAR
1 PACK OF 10 PLATES

LOT NUMBER 4433602
EXPIRY DATE 2023.12.27
PACKING DATE 2023.10.24
TEST DATE 2023.10.24
REPORTING DATE 2023.10.27

All testing in accordance with internally derived specifications, unless otherwise stated.

| Physical Characteristics | Results | Specification | Accredited Method Reference |
|--|----------|--------------------------|--|
| Appearance | Black | Black | SOP 178 Appearance and colour |
| pH (25°C) | 7.3 | 7.2 - 7.6 | SOP 53 pH |
| Fill Weight 19 ± 2 Grams | Conforms | Conforms | SOP 74 Fill volume weight check |
| Contamination @ 32 ± 2°C for >= 72 hours | Conforms | Within acceptable limits | SOP 167 Contamination Check at 22°C & 32°C |

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification. For inhibited organisms, the test medium must show no growth from the stated inoculum.

| Target Organism | Control c.f.u | Test c.f.u | Colonial Appearance | Colonial Appearance Specification | Accredited Method Reference |
|------------------------------------|---------------|------------|---------------------|-----------------------------------|---|
| Campylobacter jejuni ATCC®33291 | 44 | 32 | Grey cols | Grey cols | SOP 151 Fertility of Specified Target Organism(s)(Agar) |

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

G & M Procter Ltd. Certificate of Analysis

| Target Organism | Control c.f.u | Test c.f.u | Colonial Appearance | Colonial Appearance Specification | Accredited Method Reference |
|-------------------------------------|------------------|------------|---------------------|-----------------------------------|---|
| Campylobacter jejuni ATCC®29428 | 77 | 42 | Grey cols | Grey cols | SOP 151 Fertility of Specified Target Organism(s)(Agar) |
| Inhibited Organism | Control(cfu) | Test | Specification | Accredited Method Reference | |
| Escherichia coli ATCC®25922 | 10,000 - 100,000 | No growth | No growth | SOP 155 Inhibition | |
| Staphylococcus aureus ATCC®25923 | 10,000 - 100,000 | No growth | No growth | SOP 155 Inhibition | |

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2023.10.27

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball
Quality Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX