

CERTIFICATE OF ANALYSIS

PRODUCT PO5098A BRILLIANCE (TM) SALMONELLA
LOT NUMBER 6273805
EXPIRY DATE 2025.08.11
TEST DATE 2025.05.20
REPORTING DATE 2025.05.23

General Characteristics	Results	Specification
Colour	Conforms	Cream to pink tinged
Appearance	Conforms	Opaque
pH	7.3	7.2 -7.4
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Salmonella Typhimurium ATCC®14028	52	61	2-3mm purple/ pink colonies
Salmonella arizonae ATCC®13314	82	75	1-2mm purple / pink colonies
Escherichia coli ATCC®25922	1E+04 - 1E+05	Conforms	No growth or 1-2 mm cream colonies
Pseudomonas aeruginosa ATCC®27853	1E+04 - 1E+05	0	Complete inhibition (<= 10 cfu)

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.