

Certificate of Quality

Product Name: Cornmeal Agar w/PS80 (26mL) 10/PK
Lot Number: 221291


Product Number: R01354
Expiration Date: 2025-04-04
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Candida albicans	10231	Hyphae and Chlamydo spores Produced
Cryptococcus neoformans	34877	No Hyphae or Chlamydo spores Produced

Physical Characteristics:	
Appearance:	Light White, Clear Agar w/ Particulates
pH @ 20-25°C:	6.0 +/- 0.2

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Senior Quality Assurance Engineer