

Certificate of Quality

Product Name: KBE (Kanamycin Bile Esculin) Agar 10/PK
Lot Number: 269049


Product Number: R01510
Expiration Date: 2025-07-16
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides thetaiotaomicron	29148	Good Growth, Blackening
Bacteroides fragilis	25285	Good Growth, Blackening
Clostridium perfringens	13124	No Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth

Physical Characteristics:
Appearance: Beige-Green, Clear Agar
pH @ 20-25°C: N/A

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Senior Quality Assurance Engineer