



Certificate of Quality

Product Name: Simmons Citrate Agar (slant) 100/PK  
Lot Number: 150123


Product Number: R060494  
Expiration Date: 2025-07-16  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Enterobacter aerogenes	13048	(+) Blue-Green
Escherichia coli	25922	(-) Green Slant

Physical Characteristics:	
Appearance:	Forest Green, Clear Agar Slant
pH @ 20-25°C:	6.9 +/- 0.2

Microbial Load Evaluation:  
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed  
  
Senior Quality Assurance Engineer