

Certificate of Quality

**Product Name:** Simmons Citrate Agar (slant) 100/PK  
**Lot Number:** 218191


**Product Number:** R060494  
**Expiration Date:** 2025-12-31  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge     | ATCC  | Performance Criteria |
|------------------------|-------|----------------------|
| Enterobacter aerogenes | 13048 | (+) Blue-Green       |
| Escherichia coli       | 25922 | (-) Green Slant      |

| Physical Characteristics: |                                |
|---------------------------|--------------------------------|
| Appearance:               | Forest Green, Clear Agar Slant |
| pH @ 20-25°C:             | 6.9 +/- 0.2                    |

**Microbial Load Evaluation:**  
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed  
  
Senior Quality Assurance Engineer