

Certificate of Analysis

Product Name: Salmonella H (d) Antisera 2ml/VL
Product Code: R30160501
Lot Number: 6118705
Expiration Date: 30.09.2027
Date of Manufacture: 13.09.2024

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74AZD08Q	Salmonella D-H Antisera Filled Bulk	6115517	30.09.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Identity Slide Agglutination	Saline control must not show agglutination	Conforms
	Homologous controls must show agglutination equal to or greater than +	Conforms
	Serum under test must show agglutination equal to or greater than + with the homologous culture	Conforms
	Serum under test must not give agglutination with the heterologous culture	Conforms



Fay Curd
 Quality Assurance Manager