

Certificate of Analysis

Product Name: Salmonella H (1,7) Antisera 2ml/VL
Product Code: R30163601
Lot Number: 2114799
Expiration Date: 2020-03-31
Date of Manufacture: 2017-03-31

*Date format is DD.MM.YYYY

| Item | Description | Lot/Serial | Expire |
|-----------|------------------------------|------------|------------|
| W74AZD39V | Salmonella 1,7-H Filled Bulk | 2113030 | 2020-03-31 |

Test results carried out according to the Quality Assurance Product Release and Test Specification:

| Test | Anticipated result | Actual Result |
|--|---|---------------|
| Titre-Tube Agglutination | Serum must show agglutination to the titre stated on the bottle label plus or minus one doubling dilution | Conforms |
| | Saline control tube must remain turbid and not show agglutination | Conforms |
| Abbreviated Identity Slide Agglutination | Saline control must not show agglutination | Conforms |
| | Homologous controls must show agglutination equal to or greater than + | Conforms |
| | Serum under test must show agglutination equal to or greater than + with the homologous culture | Conforms |
| | Serum under test must not give agglutination with the heterologous culture | Conforms |



Fay Curd
Quality Assurance Manager