

## Certificate of Analysis

**Product Name:** Salmonella H Poly Phases 1&2-2ml/VL  
**Product Code:** R30858501  
**Lot Number:** 6222742  
**Expiration Date:** 31.12.2027  
**Date of Manufacture:** 03.04.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZD01E	Salmonella Poly H 1-2 Filled Bulk	6179300	31.12.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Slide Agglutination Test	Saline control must not show agglutination	Conforms
	Homologous controls must show agglutination equal to or greater than +	Conforms
	Serum under test must show agglutination equal to or greater than + with the homologous culture	Conforms
	Serum under test must not show agglutination with the heterologous culture	Conforms



Fay Curd  
 Quality Assurance Manager