

Certificate of Analysis

Product Name: Salmonella O Factor 9 (Gp D) 2ml/VL
Product Code: R30957301
Lot Number: 6233858
Expiration Date: 10.02.2028
Date of Manufacture: 01.04.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZC17W	Salmonella 9-O Antisera Filled Bulk	6192759	10.02.2028

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Identity Slide Agglutination	Saline control should give a negative reaction	Conforms
	Heterologous strain must give a negative reaction (trace is acceptable)	Conforms
	Homologous strain must give a positive reaction equal to or greater than +	Conforms



Fay Curd
 Quality Assurance Manager