

## Certificate of Analysis - Certified Reference Material

Thermo Scientific <sup>TM</sup>	<sup>I</sup> Trademark <sup>TM</sup>
Product Number	R4601241
Product Name	P. laurentii ATCC 18803 PK/5
Lot Number	206224
Usage Decision	Accepted (OK)
Expiration Date	2026-06-15

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. The results were derived from a representative sample of the batch and were obtained at the time of release. Refer to the enclosed product insert for instructions, intended use, hazard/safety requirements, and storage conditions.

Product Char Results

Purity	Demonstrates pure growth on applicable media
Viability	Recovered at acceptable level within test period
Passage	3 (Current preserved state)

Microbiological Testing	Results	Specification					
>85% Identification on Vitek 2C YST>95% Identification on MicroSEQ	99	85 - 100 95 - 100					
Microscopic Features	Pass						

These tests are performed in accordance with ISO 17025 guidelines.

Thermo Fisher Scientific has determined each loop of this reference material to be sufficiently homogeneous for its intended use.

Individual products are traceable to a recognized culture collection.

Although the Vitek(TM) panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods

Signed

Abh

Dawn Baker QA Manager

The identity, purity, and authenticity of the Licensed Products are exclusively the responsibility of Remel Inc. and not ATCC. The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative Word mark, and the ATCC Catalog Marks are trademarks of ATCC.

Remel Inc. is licensed to use these trademarks and to sell products derived from ATCC® culture.





Thermo Fisher Scientific is accredited by A2LA as a registered reference material producer certificate number 6559.01 in accordance with ISO 17034[1].

[1] ISO 17034 First Edition 2016-11-01 General requirements for the competence of reference material producers

Version 1.0

## Thermo Fisher Scientific Laboratory Report

## Isolate: 206224-1 (To be reviewed) Card Type: YST Bar Code: 2432630503579344 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Bri Yokum(BYokum)

## Bionumber: 4316755257333571 Organism Quantity:

Selected Organism: Cryptococcus laurentii

Comments:	
Comments:	

McFarland:										
Identification	Card:	YST	Lot Number:	2432630503	Expires:	Feb 7, 2025 12:00 CST				
Information	Status:	Final	Analysis Time:	17.78 hours	Completed:	Dec 27, 2024 05:09 CST				
Organism Origin	VITEK 2									
Selected Organism	99% Probability Cryptococcus laurentii									
Selected Organism	Bionumbe	er: 43167552	57333571	Confidence: Excellent identification						
Analysis Organisms and	I Tests to Se	eparate:								

Analysis Messages:

**Contraindicating Typical Biopattern(s)** 

Biochemical Details																	
3	LysA	-	4	1MLTa	-	5	LeuA	+	7	ARG	+	10	ERYa	+	12	GLYLa	-
13	TyrA	+	14	BNAG	-	15	ARBa	-	18	AMYa	-	19	dGALa	+	20	GENa	+
21	dGLUa	+	23	LACa	+	24	MAdGa	+	26	dCELa	+	27	GGT	-	28	dMALa	+
29	dRAFa	+	30	NAGA1	-	32	dMNEa	+	33	dMELa	-	34	dMLZa	+	38	ISBEa	-
39	lRHAa	+	40	XLTa	-	42	dSORa	+	44	SACa	+	45	URE	+	46	AGLU	+
47	dTURa	+	48	dTREa	+	49	NO3a	-	51	lARAa	+	52	dGATa	+	53	ESC	-
54	lGLTa	+	55	dXYLa	+	56	LATa	-	58	ACEa	+	59	CITa	-	60	GRTas	+
61	lPROa	+	62	2KGa	+	63	NAGa	+	64	dGNTa	+						

bioMérieux Customer: 1011208 System #: 504538 Thermo Fisher Scientific Laboratory Report

Autoprint Report Version: 1 of 1

Isolate: 206224-1 (To be reviewed) Card Type: YST Bar Code: 2432630503579344 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Bri Yokum(BYokum)

Bionumber: 4316755257333571 Organism Quantity:

Selected Organism: Cryptococcus laurentii

Action

Name (User ID)

Date/Time

Comment

Installed VITEK 2 Systems Version: 9.02 MIC Interpretation Guideline: AES Parameter Set Name:

Therapeutic Interpretation Guideline: AES Parameter Last Modified: Page 2 of 2