

Certificate of Quality

Product Name: P. laurentii ATCC 66036 PK/5
Lot Number: 613932

Product Number: R4601711
Expiration Date: 2009-06-30
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to performance criteria for this product.

Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description.
Traditional staining is performed.

Biochemical Analysis:

Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

CFU/loop: >10(4)

Passage: 3

Gram Reaction: Gram Positive Yeast

Identification Profile: MicroSEQ® and Vitek® 2

Appearance: Preserved Gel Matrix suspended in inoculating loop

pH: N/A

Signed



Quality Control Supervisor