

## Certificate of Analysis - Certified Reference Material

Thermo Scientific <sup>TM</sup>	Trademark <sup>TM</sup>
Product Number	R4609385
Product Name	K. pneumoniae ATCC BAA-1706 PK/5
Lot Number	117639
Usage Decision	Accepted (OK)
Expiration Date	2025-11-13

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. The results were derived from a representative sample of the batch and were obtained at the time of release. Refer to the enclosed product insert for instructions, intended use, hazard/safety requirements, and storage conditions.

Product Char Results

Purity	Demonstrates pure growth on applicable media
Viability	Recovered at acceptable level within test period
Passage	3 (Current preserved state)

Microbiological testing	Results	Specification
>85% Identification on Vitek 2C GN >95% Identification on MicroSEQ	99	85 - 100 95 - 100
Microscopic Features	Pass	

These tests are performed in accordance with ISO 17025 guidelines.

Thermo Fisher Scientific has determined each loop of this reference material to be sufficiently homogeneous for its intended use.

Individual products are traceable to a recognized culture collection.

Although the Vitek(TM) panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods

Signed

Abh

Dawn Baker QA Manager

The identity, purity, and authenticity of the Licensed Products are exclusively the responsibility of Remel Inc. and not ATCC. The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative Word mark, and the ATCC Catalog Marks are trademarks of ATCC.

Remel Inc. is licensed to use these trademarks and to sell products derived from ATCC® culture.





Thermo Fisher Scientific is accredited by A2LA as a registered reference material producer certificate number 6559.01 in accordance with ISO 17034[1].

[1] ISO 17034 First Edition 2016-11-01 General requirements for the competence of reference material producers

Version 1.0

Thermo Fisher Scientific Laboratory Report

## Isolate: 117639-1 (To be reviewed) Card Type: GN Bar Code: 2412449103351759 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Bri Yokum(BYokum)

## Bionumber: 2607734251164010

Organism Quantity:	Selected Organism: Klebsiella pneumoniae ssp pneumoniae	
Commenter		
Comments:		
		,

McFarland:											
Identification	Card:	GN	Lot Number:	2412449103	Expires:	Aug 10, 2024 13:00 CDT					
Information	Status:	Final	Analysis Time:	3.87 hours	Completed:	May 21, 2024 18:10 CDT					
Organism Origin	VITEK 2										
Selected Organism	99% Prob	99% Probability Klebsiella pneumoniae ssp pneumoniae									
	Bionumbe	er: 26077342	51164010	Confidence: Excellent identification							
Analysis Organisms and Tests to Separate:											
Analysis Messages:											

Contraindicating Typical Biopattern(s)

Biochemical Details																	
2	APPA	-	3	ADO	+	4	PyrA	-	5	lARL	-	7	dCEL	+	9	BGAL	+
10	H2S	-	11	BNAG	-	12	AGLTp	-	13	dGLU	+	14	GGT	+	15	OFF	+
17	BGLU	+	18	dMAL	+	19	dMAN	+	20	dMNE	+	21	BXYL	+	22	BAlap	-
23	ProA	-	26	LIP	-	27	PLE	+	29	TyrA	-	31	URE	+	32	dSOR	-
33	SAC	+	34	dTAG	-	35	dTRE	+	36	CIT	+	37	MNT	-	39	5KG	-
40	lLATk	+	41	AGLU	-	42	SUCT	-	43	NAGA	-	44	AGAL	+	45	PHOS	+
46	GlyA	-	47	ODC	-	48	LDC	+	53	lHISa	-	56	CMT	-	57	BGUR	-
58	O129R	+	59	GGAA	-	61	lMLTa	-	62	ELLM	-	64	lLATa	-			

bioMérieux Customer: 1011208 System #: 504538

## Thermo Fisher Scientific Laboratory Report

Autoprint Report Version: 1 of 1

Isolate: 117639-1 (To be reviewed) Card Type: GN Bar Code: 2412449103351759 Testing Instrument: 00000B4D4CA6 (01783) Setup Technologist: Bri Yokum(BYokum)

Bionumber: 2607734251164010 Organism Quantity:

Selected Organism: Klebsiella pneumoniae ssp pneumoniae

Action

Name (User ID)

Date/Time

Comment

Installed VITEK 2 Systems Version: 9.02 MIC Interpretation Guideline: AES Parameter Set Name:

Therapeutic Interpretation Guideline: AES Parameter Last Modified: