

## Certificate of Quality

**Product Name:** APINH QC Set/3  
**Lot Number:** 826546

**Product Number:** R4653070  
**Expiration Date:** 2011-04-30  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to performance criteria for this product.

### Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

### Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

### Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description.  
Traditional staining is performed.

### Biochemical Analysis:

Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

		<b>Gram Reaction:</b>	<b>Profile:</b>
Neisseria gonorrhoeae	31426	Gram Negative Diplococci	Vitek NHI
Haemophilus paraphrophilus	49917	Gram Negative Rod	Vitek NHI
Haemophilus influenzae Type B	10211	Gram Negative Rod	Vitek NHI

CFU/loop: >10(4)

Passage: Kit - All <5

Appearance: Preserved Gel Matrix suspended in loop.

pH: N/A

Signed



Quality Control Supervisor