

Certificate of Analysis - Accredited Reference Material



Thermo ScientificTM TrademarkTM

Product Number

R4717030

Product Name

E. faecalis ATCC 29212 Q+ KT/100 TST

Lot Number

247548

Usage Decision

Accepted (OK)

Expiration Date

2026-09-24

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. The results were derived from a representative sample of the batch and were obtained at the time of release. Refer to the enclosed product insert for instructions, intended use, hazard/safety requirements, and storage conditions.


Product Char Results

Purity	Demonstrates pure growth on applicable media
Viability	Recovered at acceptable level within test period
Passage	2 (Current preserved state)
Delivery	<100 CFU/0.1 mL

E. faecalis ATCC 29212 Q+ KT/100 TST	Results	Specification
>85% Identification on Vitek 2C GP	99	85 - 100
>95% Identification on MicroSEQ		95 - 100
CFU Mean	71	-
Microscopic Features	Pass	

These tests are performed in accordance with ISO 17025 guidelines. Thermo Fisher Scientific has determined each loop of this reference material to be sufficiently homogeneous for its intended use. Individual products are traceable to a recognized culture collection. Although the Vitek(TM) panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods

Signed



Dawn Baker
QA Manager

The identity, purity, and authenticity of the Licensed Products are exclusively the responsibility of Remel Inc. and not ATCC. The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative Word mark, and the ATCC Catalog Marks are trademarks of ATCC.

Remel Inc. is licensed to use these trademarks and to sell products derived from ATCC® culture.



CERT #6559.01

Thermo Fisher Scientific is accredited by A2LA as a registered reference material producer certificate number 6559.01 in accordance with ISO 17034[1].

[1] ISO 17034 First Edition 2016-11-01 General requirements for the competence of reference material producers

Version 1.0

bioMérieux Customer: 1011208
System #: 504538

Thermo Fisher Scientific
Laboratory Report

Autoprint
Report Version: 1 of 1

Isolate: 247548-1 (To be reviewed)
Card Type: GP Bar Code: 2423129403243473 Testing Instrument: 00000B4D4CA6 (01783)
Setup Technologist: Bri Yokum(BYokum)

Bionumber: 154002721753471

Organism Quantity:

Selected Organism: Enterococcus faecalis

Comments:	

McFarland:

Identification Information	Card:	GP	Lot Number:	2423129403	Expires:	Jun 21, 2026 13:00 CDT
	Status:	Final	Analysis Time:	2.90 hours	Completed:	Apr 1, 2025 15:56 CDT
Organism Origin	VITEK 2					
Selected Organism	99% Probability Enterococcus faecalis					
	Bionumber: 154002721753471 Confidence: Excellent identification					
Analysis Organisms and Tests to Separate:						
Analysis Messages:						
Contraindicating Typical Biopattern(s)						

Biochemical Details																	
2	AMY	+	4	PIPLC	-	5	dXYL	-	8	ADH1	+	9	BGAL	-	11	AGLU	+
13	APPA	-	14	CDEX	-	15	AspA	+	16	BGAR	-	17	AMAN	-	19	PHOS	-
20	LeuA	-	23	ProA	-	24	BGURr	-	25	AGAL	-	26	PyrA	+	27	BGUR	-
28	AlaA	+	29	TyrA	+	30	dSOR	+	31	URE	-	32	POLYB	+	37	dGAL	-
38	dRIB	+	39	ILATk	-	42	LAC	-	44	NAG	+	45	dMAL	+	46	BACI	+
47	NOVO	+	50	NC6.5	-	52	dMAN	+	53	dMNE	+	54	MBdG	+	56	PUL	-
57	dRAF	-	58	O129R	-	59	SAL	+	60	SAC	+	62	dTRE	+	63	ADH2s	+
64	OPTO	+															

bioMérieux Customer: 1011208
System #: 504538

Thermo Fisher Scientific
Laboratory Report

Autoprint
Report Version: 1 of 1

Isolate: 247548-1 (To be reviewed)
Card Type: GP Bar Code: 2423129403243473 Testing Instrument: 00000B4D4CA6 (01783)
Setup Technologist: Bri Yokum(BYokum)

Bionumber: 154002721753471

Organism Quantity:

Selected Organism: Enterococcus faecalis

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------