

CERTIFICATE OF ANALYSIS

PRODUCT SR0030J
POTASSIUM TELLURITE 3.5%

LOT NUMBER 6232655

EXPIRY DATE 2027.02.14

DATE OF MANUFACTURE 2025.02.14

Delivery/Customer information

Date Printed

2025.03.05

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Clear, colourless liquid	Clear, colourless liquid

Microbiological Performance	Control cfu	Test cfu	Recovery Test % Description
-----------------------------	----------------	-------------	--------------------------------

Tested in Hoyle Medium Base CM0083 enriched with 5% v/v laked horse blood

Aerobic incubation at 37°C for 36 hours

<i>Coryn. diphtheriae gravis</i> ATCC®19409	1E+04-1E+06	Dull grey/black cols
<i>Coryn. diphtheriae mitis</i> NCTC10356	1E+04-1E+06	Shiny grey/black cols
<i>C. diphtheriae intermedius</i> ATCC®14779	1E+04-1E+06	Shiny grey/black cols, dark centre

Control Medium: Columbia Blood Agar Base enriched with 5% v/v horse blood

Corynebacterium species shall produce colonies with a positive diagnostic reaction from an inoculum of 1E+04 to 1E+06 colony-forming units.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.03.05

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.