

CERTIFICATE OF ANALYSIS

PRODUCT SR0037H
POTASSIUM LACTATE 50%

LOT NUMBER 6238250

EXPIRY DATE 2027.03.04

DATE OF MANUFACTURE 2025.03.04

Delivery/Customer information

Date Printed

2025.03.18

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Clear, colourless liquid	Clear, colourless liquid

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.03.17

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com