

Delivery/Customer information

CERTIFICATE OF ANALYSIS

PRODUCT	SR0070E STAPHSTREP.SELECT.SUPP.	Date Printed 2023.09.20 Delivery No.
LOT NUMBER	3730902	Customer
EXPIRY DATE	2026.09.30	Customer Order Number
DATE OF MANUFACTURE	2023.09.11	

Physical Characteristics	Results	Specification	
Appearance	White pellet	White pellet	
Microbiological Performance	Control Test cfu cfu	Recovery Test % Description	

Tested in Columbia Blood Agar Base CM0331 enriched with 5% v/v horse blood

Aerobic incubation at 37°C for 18 hours

neubución uter e los se mous				
Streptococcus pyogenes ATCC®19615	51	44	86	Colourless/white cols, beta haem
Streptococcus pneumoniae ATCC®6303	12	10	83	Grey/white cols, alpha haem
Enterococcus faecalis ATCC®19433	80	85	106	Colourless/white cols
Enterococcus durans ATCC®19432	11	13	118	Colourless/white cols
Staphylococcus aureus ATCC®25923	10	10	100	White/orange cols
Staphylococcus epidermidis ATCC®14990	40	37	93	White/orange cols
Bacillus cereus ATCC®10876	>1E+04	0	-	No growth
Escherichia coli ATCC®25922	>1E+04	0	-	No growth
Proteus mirabilis ATCC®12453	>1E+04	0	-	No growth
Pseudomonas aeruginosa ATCC®27853	>1E+04	0	-	No growth

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu. Refer to product specification for full details.





CERTIFICATE OF ANALYSIS

PRODUCT	SR0070E
	STAPHSTREP.SELECT.SUPP.
LOT NUMBER	3730902

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release. Lot Accepted. 2023.09.20

Alarhen

.....

Carissa Courtney Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

