



CERTIFICATE OF ANALYSIS

PRODUCT SR0083A
BRUCELLA SELECTIVE SUPP.1 X 10 VIALS

LOT NUMBER 3517625

EXPIRY DATE 2024.07.31

DATE OF MANUFACTURE 2022.07.11

Delivery/Customer information

Date Printed

2022.08.30

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results		Specification	
Appearance	White pellet		White pellet	
Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Tested in Brucella Medium Base CM0169 with the addition of 10% v/v inactivated horse serum and 5% v/v of a filter-sterilised 10% w/v solution of glucose				
Incubation at 37°C for 48 hours in 10% CO₂				
<i>Staphylococcus aureus</i> ATCC®25923	>1E+04	0	-	No growth
<i>Escherichia coli</i> ATCC®25922	>1E+04	0	-	No growth
	Control cfu	Test cfu	Log(10) Reduction	
<i>Pseudomonas aeruginosa</i> ATCC®27853	3.0E+02	1.5E+01	1	Straw cols
<i>Candida albicans</i> ATCC®10231	3.5E+03	0.0E+00	3	No growth

Control Media: Tryptone Soya Agar, Sabouraud Dextrose Agar or Brucella Medium Base enriched with 5% v/v inactivated horse serum, where appropriate

Negative strains are inhibited or shall produce at least a 1 log(10) reduction when compared to the control medium.

For *Brucella abortus* Biotype 2, a satisfactory result is represented by growth characteristics comparable to the standard after incubation at 37°C for 10 days in a CO₂ atmosphere.

Refer to product specification for full details.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT SR0083A
BRUCELLA SELECTIVE SUPP.1 X 10 VIALS
LOT NUMBER 3517625

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2022.08.30

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com