

**CERTIFICATE OF ANALYSIS**

PRODUCT SR0113E
BROAD SPECTRUM BETA-LACTAMASE MIX 1X10VL

LOT NUMBER 3707792

EXPIRY DATE 2025.05.26

DATE OF MANUFACTURE 2023.08.04

Delivery/Customer information

Date Printed

2023.08.22

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Potency from certificate β -lactamase I	1105	>500
Potency from certificate β -lactamase II	71	>50
Sterility test from certificate	Satisfactory	Satisfactory

Refer to product specification for full details.

A satisfactory sterility test result is represented by no detectable growth after incubation for 14 days at 20-25°C in Tryptone Soya Broth and at 30-35°C in Fluid Thioglycollate Medium.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.08.22

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096