

# CERTIFICATE OF ANALYSIS

**PRODUCT** SR0126E  
**STREPTOCOCCUS SELECTIVE SUPPL.1X10 VIALS**

**LOT NUMBER** 6263406

**EXPIRY DATE** 2028.04.27

**DATE OF MANUFACTURE** 2025.04.28

## Delivery/Customer information

Date Printed

2025.05.21

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	White pellet	White pellet

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

## Tested with Columbia Blood Agar Base CM0331 enriched with 5% v/v horse blood

### Aerobic incubation at 37°C for 18 hours

<i>Streptococcus pyogenes</i> ATCC®19615	16	23	144	Colourless/white cols, β haem
<i>Streptococcus species Group C</i> ATCC®12449	44	37	84	Colourless/grey cols, β haem
<i>Streptococcus agalactiae</i> ATCC®13813	10	12	120	Colourless/grey cols
<i>Enterococcus faecalis</i> ATCC®19433	20	21	105	Grey cols
<i>Escherichia coli</i> ATCC®25922	>1E+04	0	-	No growth
<i>Pseudomonas aeruginosa</i> ATCC®27853	>1E+04	0	-	No growth
<i>Staphylococcus aureus</i> ATCC®25923	>1E+04	0	-	No growth

Control Medium: Columbia Blood Agar Base enriched with 5% v/v horse blood

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu. Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.21

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.