

**CERTIFICATE OF ANALYSIS**

PRODUCT SR0148A
T.T.C.SOLUTION 1 X 10 VIALS

LOT NUMBER 3636223

EXPIRY DATE 2025.03.31

DATE OF MANUFACTURE 2023.03.10

Delivery/Customer information

Date Printed

2023.04.04

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results		Specification	
Appearance	Colourless liquid		Colourless liquid	
Microbiological Performance	Control cfu	Test cfu	Recovery %	Test Description
Tested in Tergitol-7 Agar CM0793				
Aerobic incubation at 37°C for 18 hours				
<i>Escherichia coli</i> ATCC®25922	66	52	79	Yellow/orange cols & yellow zone
<i>Citrobacter freundii</i> ATCC®8090	39	40	103	Claret cols with yellow zone
<i>Enterobacter aerogenes</i> ATCC®13048	73	58	79	Claret cols, green medium
<i>Klebsiella pneumoniae</i> ATCC®29665	32	29	91	Mucoid, claret/yellow cols
<i>Shigella flexneri</i> ATCC®12022	71	67	94	Claret cols with bluish zone
<i>Enterococcus faecalis</i> ATCC®19433	>1E+04	0	-	No growth
<i>Staphylococcus aureus</i> ATCC®25923	>1E+04	0	-	Colourless cols
Aerobic incubation at 44°C for 18 hours				
<i>Escherichia coli</i> ATCC®25922	63	61	97	Yellow/orange cols & yellow zone

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu. Refer to product specification for full details.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT SR0148A
LOT NUMBER T.T.C.SOLUTION 1 X 10 VIALS
3636223

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.04.04

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com