

CERTIFICATE OF ANALYSIS

PRODUCT SR0156E
FRASER SUPPLEMENT 1 X 10 VIALS

LOT NUMBER 6203047

EXPIRY DATE 2026.12.05

DATE OF MANUFACTURE 2024.12.05

Delivery/Customer information

Date Printed

2024.12.23

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Orange/green pellet	Orange/green pellet

Microbiological Performance	Initial cfu/ml	Final cfu/ml	Log(10) Increase
-----------------------------	-------------------	-----------------	------------------

Tested with the addition of Fraser Supplement SR0156

Aerobic incubation at 37 ± 2°C for 24 ± 2 hours

<i>Listeria monocytogenes</i> ATCC®7644	5.1E+00	2.2E+07	6
<i>Listeria monocytogenes</i> ATCC®13932	6.2E+00	1.6E+07	6

Aerobic incubation at 37 ± 2°C for 48 ± 2 hours

	Initial cfu/ml	Test Result
<i>Bacillus cereus</i> ATCC®10876	2.2E+03	No aesculin hydrolysis (no blackening)

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at 37 ± 2°C for 24 ± 2 hours

Mixed cultures

Challenged with <1E+01 cfu/ml of *Listeria monocytogenes* and >1E+02 cfu/ml of *Escherichia coli* and *Enterococcus faecalis*.
A satisfactory result is represented by recovery of >10 cfu of *Listeria monocytogenes* on Brilliance Listeria Agar (ISO).
SR0257 and SR0258) using the diminishing sweep technique.

	Initial cfu/ml	Test Result
<i>L. monocytogenes</i> ATCC®13932 WDCM00021	6.2E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®8739 WDCM00012	4.6E+02	No growth
+ <i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+02	No growth
<i>L. monocytogenes</i> ATCC®13932 WDCM00021	6.2E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®25922 WDCM00013	4.4E+02	No growth
+ <i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+02	No growth
<i>L. monocytogenes</i> ATCC®13932 WDCM00021	6.2E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®8739 WDCM00012	4.6E+02	No growth
+ <i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+02	No growth
<i>L. monocytogenes</i> ATCC®13932 WDCM00021	6.2E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®25922 WDCM00013	4.4E+02	No growth
+ <i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+02	No growth
<i>L. monocytogenes</i> ATCC®35152 WDCM00109	4.7E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®8739 WDCM00012	4.6E+02	No growth

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT SR0156E
FRASER SUPPLEMENT 1 X 10 VIALS
LOT NUMBER 6203047

+ <i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+02	No growth
<i>L. monocytogenes</i> ATCC®35152 WDCM00109	4.7E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®25922 WDCM00013	4.4E+02	No growth
+ <i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+02	No growth
<i>L. monocytogenes</i> ATCC®35152 WDCM00109	4.7E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®8739 WDCM00012	4.6E+02	No growth
+ <i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+02	No growth
<i>L. monocytogenes</i> ATCC®35152 WDCM00109	4.7E+00	Blue-green cols with halo
+ <i>E. coli</i> ATCC®25922 WDCM00013	4.4E+02	No growth
+ <i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+02	No growth

SR0257 and SR0258) using the diminishing sweep technique.

Pure cultures

Growth on Brilliance Listeria Agar (ISO)

A satisfactory result is represented by no growth of *Escherichia coli* and *Enterococcus faecalis*.

<i>E. coli</i> ATCC®8739 WDCM00012	4.6E+03	No growth
<i>E. coli</i> ATCC®25922 WDCM00013	4.4E+03	No growth
<i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+03	No growth
<i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+03	No growth

Growth on Tryptone Soya Agar

A satisfactory result is represented by less than 100 cfu of *Escherichia coli* and *Enterococcus faecalis*.

<i>E. coli</i> ATCC®8739 WDCM00012	4.6E+03	Cream cols
<i>E. coli</i> ATCC®25922 WDCM00013	4.4E+03	Cream cols
<i>E. faecalis</i> ATCC®19433 WDCM00009	2.0E+03	Straw cols
<i>E. faecalis</i> ATCC®29212 WDCM00087	2.8E+03	Straw cols

Control Media: Chromogenic Listeria Agar (ISO), Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

Positive strains shall produce aesculin hydrolysis after 48 hours. Negative strains shall produce no aesculin hydrolysis after 48 hours.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2024.12.23

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.