

CERTIFICATE OF ANALYSIS

PRODUCT SR0232E
CAMPYLOBACTER GROWTH SUPPLEMENT(LIQUID)

LOT NUMBER 6215259

EXPIRY DATE 2027.01.15

DATE OF MANUFACTURE 2025.01.15

Delivery/Customer information

Date Printed

2025.02.04

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
--------------------------	---------	---------------

Colour	Dark yellow liquid	Dark yellow liquid
--------	--------------------	--------------------

Microbiological Performance

Control
cfu

Test
cfu

Recovery Test
% Description

Tested in Columbia Blood Agar Base CM0331 enriched with 7% v/v lysed horse blood

Microaerophilic incubation at 37°C for 48 hours

<i>Campylobacter jejuni</i> ATCC®29428	32	27	84	Grey/brown cols with red medium
<i>Campylobacter lari</i> ATCC®35221	15	15	100	Grey/brown cols with red medium
<i>Campylobacter coli</i> ATCC®43478	21	19	90	Grey/brown cols with red medium
<i>Campylobacter fetus</i> NCTC10348	34	44	129	Grey/brown cols without red medium

Microaerophilic incubation at 42°C for 48 hours

<i>Campylobacter jejuni</i> ATCC®29428	35	28	80	Grey/brown cols with red medium
<i>Campylobacter lari</i> ATCC®35221	18	21	117	Grey/brown cols with red medium
<i>Campylobacter coli</i> ATCC®43478	40	45	113	Grey/brown cols with red medium

Control Medium: Columbia Blood Agar Base enriched with 7% v/v laked horse blood and Campylobacter Growth Supplement SR0232

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu).
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.02.04

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.