

Certificate of Analysis

FitBag® UVM 1**PRODUCT IDENTIFICATION**

Culture medium	UVM 1
Product code (REF)	DF0863B
Product name	FitBag® UVM 1
Packaging	20 FitBag® x 4.500 ml
Expiration date (MM-GG-AAAA)	09-26-2019
Manufacturing date (MM-GG-AAAA)	09-26-2017
Lot	9201700849
Sterilization	Gamma Irradiation dose min. 27 kGy max35 kGy

CHARACTERISTICS AND MICROBIOLOGICAL PERFORMANCE

Where applicable, controlled according to ISO 11133:2014.

TESTED PARAMETERS	EXPECTED RESULTS	OBSERVED RESULTS
Appearance (powder)	Beige	Conform
Appearance (reconstitute media)	Bright Yellow	Conform
Weight (powder)	244,8 g ± 4,9 g	Conform
pH (reconstitute media)	7,4 ± 0,2 at 25 °C	Conform
Microbiological status	3 days at 30°C, no growth	Conform
	3 days at 37°C, no growth	Conform
Productivity		
Listeria monocytogenes WDCM 00021 +Escherichia coli WDCM 00013 +Enterococcus faecalis WDCM 00009	>10 colonies on ALOA. Blue green colonies with opaque halo.	Conform
Listeria monocytogenes WDCM 00109 +Escherichia coli WDCM 00013 +Enterococcus faecalis WDCM 00009		Conform
Selectivity		
Escherichia coli WDCM 00013	Total inhibition (0) on TSA	Conform
Enterococcus faecalis WDCM 00009	< 100 colonies on TSA	Conform

Lot accepted, 10-30-2017

Signed Quality Assurance Manager,

