



# Certificate of Analysis

## Novex® 16% Tricine Gel

Product No. EC66955BOX  
Lot No. 12120543  
Date of Manufacture 05DEC12  
Expiration Date 27FEB13

---

### Testing Conditions

Gels were run using Novex® Tricine SDS Running Buffer and unstained protein markers. The gels were electrophoresed at 125V until the dye front reached the bottom of the gel.

### Migration

Reduced Myosin migrated to  $0.110 \pm 0.050$  Rf, reduced Carbonic Anhydrase to  $0.320 \pm 0.050$  Rf, and reduced Aprotinin to  $0.575 \pm 0.050$  Rf.

### Straightness

Across the gel, the migration of a given protein did not vary more than 3% of the length of the gel.

### Curvature

In each of the two outermost lanes, the migration of a given protein did not vary more than 2% of the length of the gel.

### Appearance

Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat.

### Overall Result

PASS

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.lifetechnologies.com](http://www.lifetechnologies.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)

---

Rick Fletcher  
Director, Quality  
Issued on 07 Dec 2012