

PRODUCT CERTIFICATION

LOT NUMBER	<u>LP5357010</u>	CATALOG NUMBER	<u>114-060A/LP</u>
PROCESS DATE	<u>1/8/2026</u>	METHOD	<u>7.5.1.14 & 7.5.1.13</u>
MANUFACTURER	<u>Tricor</u>	RAW LOT #	<u>24080105</u>
RAW PART #	<u>319288</u>		
DESCRIPTION	<u>60ml Amber Boston Round</u>		

This is to certify that the item(s) as stated above have been processed in a Class 100 Cleanroom and packaged under a Class 10 hood, using Method 7.5.1.14 and 7.5.1.13. The final rinse meets current USP test requirements for Water for injection as used in manufacturing.

ANALYSIS

ANALYTE	SPECIFICATION	RESULT
Particulates	<10pcs/mL at 0.5 microns	<10pcs/mL at 0.5microns

This is to certify that this lot was tested by Thermo Fisher Scientific full service laboratory and found to comply with Thermo Fisher Scientific's specifications for this product.

Approved By: *Levi Nott*
 Levi Nott 1/12/2026
 Quality Assurance