

**PRODUCT CERTIFICATION**

LOT NUMBER DP5255020 CATALOG NUMBER 130-04C/PF  
 EXP DATE: 9/25/27 PROCESS DATE 9/25/25  
 DESCRIPTION 4oz. Clear WM w/ cap

This is to certify that the item(s) as stated above have been processed in a Class 100 Cleanroom and packaged under a Class 10 hood, using Method 7.5.1.14 or 7.5.1.99 and 7.5.1.23 (Bottle), 7.5.1.14 or 7.5.1.98 and 7.5.1.13 (Cap and Liner). The final rinse meets current USP test requirements for Water for injection as used in manufacturing. The depyrogenation oven cycle has been validated to meet 3-Log endotoxin inactivation.

Raw Material Description	Supplier/Manufacturer	Vendor Part #	Vendor Lot #
4oz CLEAR WM	BERLIN	GLA-00857/286158	1051722
58-400 PP CT CAP	BERLIN	00398	202505705048
58MM WH PTFE LINER	JENSEN	CLT1009-58C-MJ	EN07-222208-024773-RF-7, EN10-222208-023366-RF-7

**ANALYSIS**

ANALYTE	SPECIFICATION	RESULT
Endotoxin	<0.06 EU/mL LAL	<0.06 EU/mL LAL

Thermo Fisher Scientific certifies that the item(s) in above referenced shipment were processed, tested, and meets Thermo Fisher Scientific's specifications for this product.

Approved By:   
 Christina Edison 10/6/25  
 Quality Assurance