

Thermo

SCIENTIFIC

236 CARDIFF VALLEY ROAD
ROCKWOOD, TN 37854
(800) 228-4931 phone
(865) 354-3466 fax

May 4, 2015


This document certifies the sterility of the following:

Catalog Number	Lot Number	Date of Manufacture	Date of Sterilization	Date of Expiration	Minimum Dose	Maximum Dose
N411-0250	00061400	02/21/2015	03/17/2015	*03/17/2017	32.0 kGy	38.9 kGy

Specified Dose Range: 25.0 - 45.0 KGY

*When stored in a temperature and moisture controlled environment out of direct sunlight while maintaining packaging integrity, sterility is maintained for a period of two years.

Sincerely,



Ryan Cassaro
Quality Assurance Department