

## Certificate of Analysis

1 Reagent Lane  
 Fair Lawn, NJ 07410  
 201.796.7100 tel  
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System  
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	J10907	Quality Test / Release Date	11/10/2021
Lot Number	218252		
Description	AGAR, NOBLE AGAR, ULTRAPURE		
Country of Origin	Spain	Suggested Retest Date	Apr/2024
Chemical Origin	Organic - Plant		
BSE/TSE Comment	BSE/TSE Free		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Pale cream color
ABSORBANCE 1.5 % SOL AT 450NM		<= 0.150	0.068
ASH	%	<= 1.6	1.3
ELECTROENDOSMOSIS		<= 0.450	0.36
GEL POINT (1.5% GEL)	DEGREES	Inclusive Between 32 - 37.5	35.3
GEL STRENGTH (1.5% GEL)	G/CM^2	>= 700	1000
GEL pH (1.5% GEL)		Inclusive Between 5.0 - 7.0	6.4
MELTING POINT (1.5%)	DEGREES CELSIUS	Inclusive Between 80 - 95	89.0
MOISTURE	%	<= 10	5.6
RESISTIVITY	Ohms	>= 20,000	59177



Julian Burton - Quality Control Manager – Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

\*Based on suggested storage condition.