

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	A202	Quality Test / Release Date	03/04/2021
Lot Number	207726	Expiration Date	Feb/2024
Description	NITRIC ACID FUMING, 90%, REAGENT A.C.S.		
Country of Origin	United States		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, yellow, fuming liquid
ARSENIC (As)	ppm	<= 0.3	<0.3
ASSAY	%	>= 90.0	92.26
CHLORIDE	ppm	<= 0.7	<0.7
DILUTION TEST	PASS/FAIL	= PASS TEST	PASS TEST
DISSOLVED OXIDES	%	<= 0.10	<0.1
HEAVY METALS (as Pb)	ppm	<= 5	<5
IGNITION RESIDUE	%	<= 0.002	<0.002
IRON (Fe)	ppm	<= 2	<0.8
SULFATE (SO4)	ppm	<= 5	<5



Julian Burton - Quality Control Manager - Bridgewater

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.