

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	J77500	Quality Test / Release Date	07/26/2024
Lot Number	243765		
Description	TBW WITH TWEEN (TBST), 20X SOLUTION, MOLECULAR BIOLOGY GRADE, ULTRAPURE		
Country of Origin	United States	Suggested Retest Date	Jul/2029
Comment	Form: 20X TBS (Tris Buffered Saline) with Tween 20, solution consisting of 500 mM Tris, 60 mM KCl, 2.8 M NaCl and 1.0% Tween 20 in high purity DI H2O, 0.2 micron filtered.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	clear, colorless solution
DNASE (ENDO)	PASS/FAIL	= PASS TEST	PASS TEST
DNASE (EXO)	PASS/FAIL	= PASS TEST	PASS TEST
PROTEASE	PASS/FAIL	= PASS TEST	PASS TEST
pH OF 20X SOLUTION		Inclusive Between 7.3 - 7.5	7.5



Harout Sahagian - Quality Control Manager - Fair Lawn