

Certificate of Analysis

1 Reagent Lane Fair Lawn, NJ 07410 201.796.7100 tel 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	BP1338	Quality Test / Release Date	02/10/2023	
Lot Number	226121			
Description	TRIS-EDTA (TE BUFFER), 100X SOLUTION			
Country of Origin	United States	Suggested Retest Date	Nov/2025	
Chemical Origin	Organic - non animal			
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.			

N/A				
Result Name	Units	Specifications	Test Value	
APPEARANCE		REPORT	clear, colorless liquid	
CONDUCTIVITY OF SOLUTION	MICROMHOS		731.6	
DNASE	DETECTED/NOT DETECT.	= NOT DETECTED	NOT DETECTED	
PROTEASE	DETECTED/NOT DETECT.	= NOT DETECTED	NOT DETECTED	
RNASE	DETECTED/NOT DETECT.	= NOT DETECTED	NOT DETECTED	
pH OF A 1X SOLUTION		Inclusive Between 7.9 - 8.1	8.0	

Harout Sahagian - Quality Control Supervisor - Fair Lawn