

Certificate of Analysis

1 Reagent Lane Fair Lawn, NJ 07410 201.796.7100 tel 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

| Catalog Number | E139 | Quality Test / Release Date | 11/04/2019 | |
|-------------------|---|-----------------------------|------------|--|
| Lot Number | 196284 | | | |
| Description | PETROLEUM ETHER, A.C.S. | | | |
| Country of Origin | United States | Suggested Retest Date | Nov/2024 | |
| Chemical Origin | Organic - non animal | | | |
| BSE/TSE Comment | No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product. | | | |

| N/A | | | | |
|-------------------------------|-----------------|---------------------------------|---------------------------|--|
| Result Name | Units | Specifications | Test Value | |
| APPEARANCE | | REPORT | CLEAR,COLORLESS LIQUID | |
| ACIDITY | PASS/FAIL | = PASS TEST | PASS TEST | |
| BOILING RANGE (36 - 60 DEG C) | PASS/FAIL | = PASS TEST | PASS TEST | |
| BOILING TEMP (BEGINNING) | DEGREES CELSIUS | Inclusive Between 36 - 39 | 38 | |
| BOILING TEMP (ENDPOINT) | DEGREES CELSIUS | Inclusive Between 52 - 60 | 58 | |
| COLOR | APHA | <= 10 | 5 | |
| COPPER STRIP CORROSION | PASS/FAIL | = PASS TEST | PASS TEST | |
| DISTILLATION RESIDUE | PASS/FAIL | = PASS TEST | PASS TEST | |
| DOCTOR TEST | SWEET/FAIL | = PASS TEST | PASS TEST | |
| EVAPORATION RESIDUE | % | <= 0.001 | <0.0001 | |
| FRACTION DISTILLS < 40 C | % | <= 60 | <10 | |
| FRACTION DISTILLS < 54 C | % | >= 95 | 95 | |
| HEAVY OILS & FATS | PASS/FAIL | = PASS TEST | PASS TEST | |
| IDENTIFICATION | PASS/FAIL | = PASS TEST | PASS TEST | |
| SAPONIFICATION VALUE | MG KOH/GRAM | <= 0.015 | 0.013 | |
| SPECIFIC GRAVITY @ 25 DEG C | | Inclusive Between 0.630 - 0.660 | 0.64 | |
| STEARIN EVAPORATION RESIDUE | GMS. | <= 0.003 | <0.0003 | |
| UNSATURATED COMPOUNDS | PASS/FAIL | = PASS TEST | PASS TEST | |

Deffor

Peter Yang - Quality Control Manager - Bridgewater

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above. If there are any questions with this certificate, please call at (800) 227-6701.

^{*}Based on suggested storage condition.