

Certificate of Analysis

RNase AWAY™ Decontamination Reagent

Product No. 10328011, 10594063

Lot No. 21280223

Date of Manufacture 31-Jul-2021

RNase Test

All apparatus was treated with Diethyl Pyrocarbonate (DEPC) to inhibit possible RNase contamination from outside sources of the experiment. The bases of five (5) laboratory glassware were intentionally contaminated with RNase. Select pieces were treated with RNase AWAY®, after which the pieces were subjected to a combination of DI water rinses and wiping dry. The pieces were extracted with DEPC water, and the extracts were then added to an RNA standard and incubated. Following incubation, the samples were analyzed by gel electrophoresis and evaluated for degradation of the RNA.

Results: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 06-Feb-2023