

# Certificate of Analysis

## RNase AWAY™ Decontamination Reagent

Product No. 10328011, 10594063

Lot No. 21280224

Date of Manufacture 30-Jul-2021

Expiration Date 31-Jan-2024

---

### RNase Test

All apparatus was treated with Diethyl Pyrocarbonate (DEPC) to inhibit possible RNase contamination from outside sources of the experiment. The bases of five (5) laboratory glassware were intentionally contaminated with RNase. Select pieces were treated with RNase AWAY®, after which the pieces were subjected to a combination of DI water rinses and wiping dry. The pieces were extracted with DEPC water, and the extracts were then added to an RNA standard and incubated. Following incubation, the samples were analyzed by gel electrophoresis and evaluated for degradation of the RNA.

Results: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoyn Joseph  
Director, Quality  
Issued on 14-Mar-2023