

Certificate of Analysis

Colloidal Blue Staining Kit

Product No. 10593043, LC6025

Lot No. 3065270

Date of Manufacture 13-Dec-2024

Expiration Date 28-Oct-2025

This kit contains Coomassie Stain Reagent A (part number 467015) and Coomassie Stain Reagent B (part number 467016).

Assay: Stainer A Refractive Index

Specification: (33.4 - 34.8)% Brix

Result: Meets Specification

Assay: Sensitivity

Specification: 6.25 ng BSA must be visible in 7 +/- 1 hour

Result: Meets Specification

Assay: Background

Specification: Must be free from uneven staining and contaminant bands.

Result: Meets Specification

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 31-Dec-2024