

# Certificate of Analysis

## Lipofectamine™ 2000 Transfection Reagent

Product No. 11668500, 10600714

Lot No. 2903538

Date of Manufacture 28-May-2024

Expiration Date 21-May-2026

---

### Contamination

Specification: No evidence of fungal or bacterial growth after inoculation on blood agar, Sabouraud Dextrose Agar, and fluid Thioglycollate after three days.

Result: Meets specification

### Analytical Testing

Product is tested by quantitative high performance liquid chromatography and particle size measurement.

Result: Meets specification

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 29-May-2024