

# Certificate of Analysis

## Platinum™ Quantitative RT-PCR ThermoScript™ One-Step System

Product No. 11731023, 10525013

Lot No. 2817316

Date of Manufacture 09-Oct-2023

---

### Quantitative RT-PCR Assay

Specification: The correlation coefficient for the standard curve (amplification plot) covering 6 logs of input HeLa RNA concentration must be  $> 0.980$ .

Results: Meets specification.

Specification: The slope of the standard curve must be greater than  $-3.7$ .

Results: Meets specification.

Specification: No signal above the threshold observed before 39 cycles in at least 2 of 6 no-template controls.

Results: Meets specification.

Specifications: The average cycle threshold value for reactions containing 1 pg of HeLa RNA must be less than 39.

Results: Meets specifications.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoyn Joseph  
Director, Quality  
Issued on 29-Jan-2025