

# Certificate of Analysis

## SuperScript™ VILO™ Master Mix

Product No. 11755050, 12013679, 15734032

Lot No. 2945870

Date of Manufacture 02-Aug-2024

---

### SuperScript® VILO™ qRT-PCR Functional Assay

Specification: This product is tested functionally in qRT-PCR using total RNA. Kinetic analysis must demonstrate a linear dose response between -3.20 and -3.60 over six orders of magnitude from 2.5 µg – 2.5 pg with a correlation coefficient  $\geq 0.990$ .

Result: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 05-Aug-2024