

Certificate of Analysis

SuperScript® VILO™ MasterMix - 250rxn

Product No. 11755250C, 11755250, 11755-250, 12023679

Lot No. 2881785

Date of Manufacture 29-Apr-2024

SuperScript® VILO™ qRT-PCR Functional Assay

Specification: This product is tested functionally in qRT-PCR using total RNA. Kinetic analysis must demonstrate a linear dose response between -3.20 and -3.60 over six orders of magnitude from 2.5 μ g – 2.5 pg with a correlation coefficient > 0.990.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific Life Sciences Solutions 5781 Van Allen Way Carlsbad, CA, USA 92008 https://www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com

Chevohn Joseph Director, Quality Issued on 30-Apr-2024