

Certificate of Analysis

SuperScript™ VILO™ Master Mix

Product No. 11755250C, 11755250, 12023679

Lot No. 2980878

Date of Manufacture 04-Oct-2024

SuperScript® VILO™ qRT-PCR Functional Assay

Specification: This product is tested functionally in qRT-PCR using total RNA. Kinetic analysis must demonstrate a linear dose response between -3.20 and -3.60 over six orders of magnitude from 2.5 µg – 2.5 pg with a correlation coefficient ≥ 0.990 .

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 07-Oct-2024