

# Certificate of Analysis

## SYBR GreenER™ qPCR SuperMix for ABI PRISM™ Instrument

Product No. 11760100, 10512773

Lot No. 3063130

Date of Manufacture 02-Jan-2025

---

### Product Qualification

This product is functionally tested using plasmid DNA quantified by digital PCR to ensure robust amplification of DNA targets.

Specification: The slope must be  $-3.4 \pm 0.2$ .

Result: Meets specification.

Specification: The amplification efficiency must be  $\geq 90\%$ .

Result: Meets specification.

Specification: The correlation coefficient (  $R^2$  ) must be  $\geq 0.99$ .

Result: Meets specification.

Specification: Average Ct value of 1,000 copy level must be between  $26.5 \pm 1.5$ .

Result: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoyn Joseph  
Director, Quality  
Issued on 06-Jan-2025