

# Certificate of Analysis

## PureLink™ RNA Mini Kit

Product No. 12183020, 10359103

Lot No. 2896319

Date of Manufacture 26-Apr-2024

---

The PureLink™ RNA Mini Kit is functionally tested for optimum performance using the following tests:

### Gel Electrophoresis

Total RNA is analyzed by gel electrophoresis. Total RNA should appear intact with no degradation. Two ribosomal RNA bands should be present.

Result: Meets specification

### Yield

OD260/OD280 ratio must be between 1.8 - 2.2.

Result: Meets specification

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 01-May-2024