

# Certificate of Analysis

## Lipofectamine™ LTX Reagent with PLUS™ Reagent

Product No. 15338030, 12353563

Lot No. 2915253

Date of Manufacture 08-May-2024

Expiration Date 28-Nov-2025

---

### Functional Assay

CHO-K1 cells are transfected with a *lacZ* reporter plasmid. The ratio of gene expression between the test lot and a reference lot must be  $1.0 \pm 0.15$ . The peak activities of the test and reference lots must be within  $\pm 0.1 \mu\text{l}$ .

Result: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoehn Joseph  
Director, Quality  
Issued on 09-May-2024