

Certificate of Analysis

UltraPure™ Guanidine Isothiocyanate Solution

Product No. 15577018, 11598746

Lot No. 2552264

Date of Manufacture 24-Oct-2022

Guanidine Isothiocyanate Solution is functionally tested for optimum performance using the following tests:

Specifications:

1. pH Determination: The pH must be 7.20 to 7.80.
2. Conductivity Determination: The conductivity of a 1:100 dilution must be 3.90 to 4.70 mS/cm.
3. Appearance: The solution must be colorless and free of any particulates.

Results: Meets specifications.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 16-Nov-2022